

PhACTS_dx Infectious Disease ELISA

03/14/2000

Plate File: rube_01c.plt
 Test Date: 07/21/1999
 Test Time: 04:06:40

Clark

Workspace Name: w1rub1
 Results Recorded by Workstation 1
 Data Entry by Workstation 1

Test Validation for Rub IgG Master Lot 09178

Rule 1. Test Date (2000/03) <= Expiration Date (1999/09) VALID

Rule 2. Sample Pos Abs QC Req.
 Blank A01 0.008 < 0.150 VALID
 NC B01 0.011 <= 0.250 VALID
 LPC C01 0.762 >= 0.250 VALID
 LPC D01 0.775 >= 0.250 VALID
 HPC E01 1.427 >= 0.500 VALID

QC Passed

Rule 2. QC Calculation QC Req.
 HPC / Mean LPC 1.857 >= 1.200 VALID
 NC / Mean LPC 0.014 <= 0.900 VALID

Determination of Calibrator Value

Mean LP (0.769) x Factor (0.500) = Calibrator Value (0.384)

Sample ID	Loc	Result	Abs	ISR	Sample ID	Loc	Result	Abs	ISR		
+	rc910	A02	POSITIVE	0.467	1.22	+	24	A05	POSITIVE	0.454	1.18
+	1	B02	POSITIVE	0.438	1.14	+	25	B05	POSITIVE	0.488	1.27
=	2	C02	EQV	0.411	1.07	+	26	C05	POSITIVE	0.501	1.30
+	3	D02	POSITIVE	0.457	1.19	+	27	D05	POSITIVE	0.501	1.30
=	4	E02	EQV	0.416	1.08	+	28	E05	POSITIVE	0.565	1.47
+	5	F02	POSITIVE	0.436	1.13	+	29	F05	POSITIVE	0.489	1.27
+	6	G02	POSITIVE	0.441	1.15	+	30	G05	POSITIVE	0.529	1.38
=	7	H02	EQV	0.376	0.98	+	31	H05	POSITIVE	0.501	1.30
+	8	A03	POSITIVE	0.456	1.19	+	32	A06	POSITIVE	0.465	1.21
+	9	B03	POSITIVE	0.458	1.19	+	33	B06	POSITIVE	0.527	1.37
+	10	C03	POSITIVE	0.497	1.29	+	34	C06	POSITIVE	0.497	1.29
+	11	D03	POSITIVE	0.447	1.16	=	35	D06	EQV	0.404	1.05
+	12	E03	POSITIVE	0.498	1.30	+	36	E06	POSITIVE	0.458	1.19
+	13	F03	POSITIVE	0.441	1.15	+	37	F06	POSITIVE	0.465	1.21
+	14	G03	POSITIVE	0.441	1.15	+	38	G06	POSITIVE	0.527	1.37
+	15	H03	POSITIVE	0.472	1.23	+	39	H06	POSITIVE	0.459	1.19
+	16	A04	POSITIVE	0.432	1.12	+	40	A07	POSITIVE	0.455	1.18
+	17	B04	POSITIVE	0.451	1.17	=	41	B07	EQV	0.416	1.08
=	18	C04	EQV	0.413	1.07	+	42	C07	POSITIVE	0.444	1.16
=	19	D04	EQV	0.399	1.04	+	43	D07	POSITIVE	0.491	1.28
=	20	E04	EQV	0.397	1.03	+	44	E07	POSITIVE	0.496	1.29
=	21	F04	EQV	0.410	1.07	=	45	F07	EQV	0.380	0.99
+	22	G04	POSITIVE	0.432	1.12	+	46	G07	POSITIVE	0.483	1.26
=	23	H04	EQV	0.392	1.02	+	47	H07	POSITIVE	0.445	1.16

--- End of List ---